



ALL INDIA INSTITUTE OF MEDICAL SCIENCES
Department of Pediatrics, Mother & Child Block
ANSARI NAGAR, DELHI-110029

Date 27-02-2024

Subject: Proposal for procurement of ADE M-118600 Electronic Baby Scale, Qty. 7 for Neonatal Intensive Care Unit and Pediatric Intensive Care Unit at Department of Pediatrics, Mother & Child Block, AIIMS, New Delhi-110029 on PAC basis and invitation of comments/objections against such consumables if any, thereof.

The Neonatal Intensive Care Unit (NICU) and Pediatric Intensive Care Unit (PICU) at Department of Pediatrics, Mother & Child Block, AIIMS, New Delhi is in the process of procuring of ADE M-118600 Electronic Baby Scale Qty. - 7 for performing weight measurement of newborns and infants in children at NICU and PICU of at Mother & Child Block, in Dept. of Pediatrics.

Therefore, this document is being uploaded for open information to all principal manufacturers or their authorized distributor with the intention to invite comments and/or objections with regards to the proprietary nature of the aforesaid product if any.

The comments/objections if any should be received in the office of the Stores Office, Room no. 817, 8th Floor, Department of Pediatrics, Mother & Child Block, AIIMS, Ansari Nagar, New Delhi - 110029 on or before **15/03/2024** up to 05.00 PM with contact details, failing which it will be presumed that no other bidder/ manufacturer/vendor is having any comments/objection against the proposition and the case will be decided purely on merit.

The matter pertains to the approval of the competent authority within the Department of Pediatrics at AIIMS.


Dr. Rachna Seth
Professor & O/IC Stores
डॉ. रचना सेठ
Dr. RACHNA SETH
Professor
बालरोग चिकित्सा विभाग/Department of Pediatrics
अ.भा.आ.सं., नई दिल्ली/A.I.I.M.S., New Delhi-29

RM/AIIMS-DEL/048/2023-24

20.01.2024

The Head of Department
Department of Pediatrics
All India Institute of Medical Sciences,
Ansari Nagar,
New Delhi-110029.

Dar Sir,

Thank you very much for showing your interest in **ADE Electronic Baby Scale**. Please find enclosed herewith proforma for the same.

We are exclusive distributor of ADE GmbH, Germany in India and provide sales and service on their behalf..

We also enclosed copy of CE (European) and ISO certificate and copy of agency agreement, our GST Registration and proprietary certificate from ADE Germany GmbH & Co.KG, Germany for your ready reference.

Hope you will find the above in order and we will look forward to receive your valuable orders.

Thanking you,

Truly yours,


For Rustagi Medisurg



Vijay Rustagi

Please process.

Thanks


डॉ. म. जीवा संकर
Dr. M. Jeeva Sankar, MD,DM
अपर आचार्य/Additional Professor
बालरोग चिकित्सा विभाग/Department of Pediatrics
अ.भा.आ.सं. नई दिल्ली-29/A.I.I.M.S. New Delhi-29

RM/AIIMS-DEL/047/2023-24

20.01.2024

The Head of Department
Department of Pediatrics
All India Institute of Medical Sciences,
Ansari Nagar, New Delhi-110002.

Proforma Invoice

Description

ADE Germany Gmbh & Co. KG., Germany

ADE M-118600 Electronic Baby Scale

For use in Neonates, Pediatric and in wards.

The gently contoured bay tray give generous
as well as secure space for small infants.

Large, easy to read LCD display with backlit.

Automatic Zero setting & all function test

Capacity: 20kg

Graduation: 5gm

Unit Switch function Kg / lbs

Tare & Hold Function

Automatic switch off, Low Battery Indicator

Mains adaptor and batteries operation

Scale work on 4 AA normal dry batteries

CE European and ISO certified.


Supplied with operation manual in English.


Unit	New Delhi	Price
Each	25183	Rs.49,500.00

Terms and Conditions:

1. Warranty : One year from the date of installation.
2. Price : Valid for 60 days from the date of Proforma invoice.
3. Delivery : Within 90 days from the date of confirmed supply orders.
4. Taxes : GST extra 18% at the time of delivery.
5. Installation : Free of charge by our company.
6. Payment : Within 30 days from the date of satisfactory installation..

For Rustagi Medisurg


Vijay Rustagi


Dr. M. Jeeva Sankar, MD,DM
अपर आचार्य/Additional Professor
बालरोग चिकित्सा विभाग/Department of Pediatrics
अ.भा.आ.सं. नई दिल्ली-29/A.I.I.M.S. New Delhi-29

20.01.2024

The Head of Department
Department of Paediatric,
A.I.I.M.S.,
New Delhi – 110029.

Fall Clause Price Certificate

This is to certify that our company has not supplied **ADE M-118600 Electronic Baby Scale manufactured by ADE Germany Gmbh & Co. KG., Germany** at lower rates to any Government Hospital or Institute than quoted to The Head of Department, Department of Paediatric, A.I.I.M.S., New Delhi – 110029 during the past 1 year as per Fall Price Clause adhered by other Government Institutions.

For Rustagi Medisurg



Vijay Rustagi



डा. म. जेवा संकर
Dr. M. Jeeva Sankar, MD,DM
अपर आचार्य/Additional Professor
बालरोग चिकित्सा विभाग/Department of Pediatrics
अ.भा.आ.सं. नई दिल्ली-29/A.I.I.M.S. New Delhi-29

ADE Germany GmbH, Neuer Höltigbaum 15, D-22143 Hamburg

06.04.2023

PROPRIETARY CERTIFICATE
TO WHOM IT MAY CONCERN

We hereby confirm that the following ADE Medical Scales and Measuring Rods:

- ADE-M-118600 Electronic Baby Scale
- ADE-M-118600-01 Electronic Baby Scale with Stadiometer
- ADE-M-316600 Electronic Floor Scale
- ADE-M-403660 Electronic Chair Scale
- ADE - PWN5 Electronic Diaper and Nappy Scale
- ADE-MZ-10023 Mechanical Height Road
- ADE-MZ-10038 Mechanical Telescopic Stadiometer with metal base.

are proprietary item and only manufactured by ADE Germany GmbH & Co. KG, Germany. No other company or any individual authorized to manufacture the same name and specifications.

We also confirm that RUSTAGI MEDISURG, 87, Darya Ganj, New Delhi – 110002 / India is our exclusive distributor to quote, negotiate and provide service on our behalf. No other company or individual other than RUSTAGI MEDISURG is authorized to submit offers and provide sales and service of ADE products in India.


The above information is true and correct.

For ADE GERMANY GmbH & Co. KG



Authorized Signatory

For Rustagi Medisurg



Auth. Signatory

ADE Germany GmbH, Neuer Höltigbaum 15, D-22143 Hamburg

Tel +49 40 432776-0
Fax +49 40 432776-10

HypoVereinsbank
Commerzbank

IBAN: DE15 2003 0000 0000 1377 86 BIC: HYVEDEMM300
IBAN: DE10 2004 0000 0619 7875 00 BIC: COBADEFFXXX

Tax No. 43/600/01372
VAT No. DE118958935

EG-Zertifikat / EC-Certificate

gem. 93/42/EWG Anhang V / acc. 93/42/EEC Annex V
für den Aspekt der Messfunktion / for the aspect of the measuring function

Hiermit wird bescheinigt, dass die Firma / This certifies, that the company

ADE Germany GmbH
Hammer Steindamm 27-29
22089 Hamburg
Deutschland

für die Produkte / die Kategorie: Liste der Produkte siehe Anlage 1
for the products / product category: List of products see annex 1


Medizinische Waagen und Längenmessgeräte Medical scales and length measuring devices

ein Qualitätssicherungssystem für die Produktion, beschränkt auf die Herstellungsschritte in Zusammenhang mit der Konformität der Produkte mit den messtechnischen Anforderungen nach Maßgabe des Anhang V der Richtlinie 93/42/EWG, anwendet. Die Übereinstimmung mit den Anforderungen wurde in einem Audit nachgewiesen. Zusätzlich zur CE-Kennzeichnung muss die Kennnummer der Benannten Stelle angebracht werden. Die Gültigkeit dieses Zertifikats beruht auf der Aufrechterhaltung des Qualitätssicherungssystems in Übereinstimmung mit den Anforderungen der Richtlinie und seiner Überwachung durch die Benannte Stelle gem. Anhang V Abschnitt 4. Das Zertifikat ist unter keinen Umständen übertragbar.

has established a quality system for production restricted to the conformity of the devices with metrological requirements acc. to the requirements of Annex V of the directive 93/42/EEC. Conformity with the requirements was proved within an audit. Additional to the CE-marking the notification number of the Notified Body has to be affixed. The validity of this certificate is based on the maintenance of the quality system in accordance with the requirements of the directive and its surveillance by the Notified Body according Annex V section 4. The certificate may not be transferred under any circumstances.

Reg.-Nr. / Reg.-No. 78 238 056755
Bericht Nr. / Report No. 3529 1644

Gültigkeit / Validity
von / from 2021-05-05
bis / until 2024-05-26
Edition 5




Zertifizierungsstelle für Medizinprodukte
Certification body for medical devices

Essen, 2021-05-05

TÜV NORD CERT GmbH Langemarckstraße 20 45141 Essen www.tuev-nord-cert.de medical@tuev-nord.de

Benannte Stelle Kenn-Nr. 0044 / Notified Body ID. No. 0044



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-236.10.16

For Rustagi Medisurg


Auth. Signatory

TÜV NORD CERT GmbH • P.O. Box 10 32 61 • 45032 Essen • Germany

ADE Germany GmbH
Mr. Bornhofen
Neuer Höltigbaum 15
22143 Hamburg
Germany

TÜV NORD CERT GmbH

Am TÜV 1
45307 Essen
Germany

Phone: +49 201 825-0
Fax: +49 201 825-2517

Info.tncert@tuev-nord.de
www.tuev-nord-cert.com

TÜV®

Our / Your Reference
ZA:35291644 /
Reg. No 78238056755

Contact
TÜV NORD CERT Medical
E-Mail: medical@tuev-nord.de

Direct Dial
Phone: -2236

Date
13.06.2022

Written confirmation correcting (Change of address of ADE Germany GmbH) or complementing information on valid certificates according to Directive 93/42/EEC taking into account the provisions of Regulation (EU) 2017-745, Article 120.

TÜV NORD CERT GmbH, Notified Body for medical devices, identification number 0044.

In accordance with the provisions of Regulation (EU) 2017/745, Article 120, certificates issued under Directive 93/42/EEC that are still valid may no longer be changed. This letter confirms correcting or complementing information of the following valid certificate with reg. no 78238056755

Change of address of ADE Germany GmbH:

Previous address: ADE Germany GmbH, Hammer Steindamm 27-29, 22089 Hamburg / Germany

Current address ADE Germany GmbH, Neuer Höltigbaum 15, 22143 Hamburg / Germany

i.V. Kevin Mühlenberg
Head of Project Management
Medical Devices International
TÜV NORD CERT GmbH

i.A. Mandy Unger
Project Manager

Headquarters
TÜV NORD CERT GmbH
Am TÜV 1
45307 Essen
Phone: +49 201 825-0
Fax: +49 201 825-2517
info.tncert@tuev-nord.de
www.tuev-nord-cert.com

Director
Dipl.-Ing. Wolfgang Wielpütz
Dipl.-Oec. Sandra Gehartz


Registration Office
Amtsgericht Essen
HRB 9976
VAT No.: DE 811389923
Tax No.: 111/5706/2193

Deutsche Bank AG, Essen
BIC (SWIFT-Code): DEUTDE33XXX
IBAN-Code: DE 26 3607 0050 0607 8950 00

TÜV NORD GROUP



For Rustagi Medisurg


Auth. Signatory

ZERTIFIKAT

für das Managementsystem nach
DIN EN ISO 9001 : 2015

Der Nachweis der regelwerkskonformen Anwendung wurde erbracht und wird gemäß
TÜV NORD CERT-Verfahren bescheinigt für

ADE Germany GmbH & Co. KG
Hammer Steindamm 27-29
22089 Hamburg
Deutschland



mit den Standorten/Gesellschaften gemäß Anlage

Geltungsbereich

Entwicklung, Herstellung, Vertrieb und Service
WÄGETECHNIK, GROSSKÜCHENMASCHINEN
MEDIZINISCHE WAAGEN

Zertifikat-Registrier-Nr. 78 100 056755
Auditbericht-Nr. 3525 5699

Gültigkeit vorheriges Zertifikat: 2020-01-04

Gültig von 2020-01-16

Gültig bis 2023-01-04

Erstzertifizierung 2005



Zertifizierungsstelle
der TÜV NORD CERT GmbH

Essen, 2020-01-15

Diese Zertifizierung wurde gemäß TÜV NORD CERT-Verfahren zur Auditierung und Zertifizierung durchgeführt und wird
regelmäßig überwacht.

Die Gültigkeit kann unter <https://www.tuev-nord.de/de/unternehmen/zertifizierung/zertifikatsdatenbank> verifiziert werden.

TÜV NORD CERT GmbH


Langemarckstraße 20

45141 Essen

www.tuev-nord-cert.de



For Rustagi Medisurg


Auth. Signatory

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI – 110029**

PROPRIETARY/SPECIFIC BRAND GOOD CERTIFICATE

1. Item/Type/Model no. required along with specification	ADE Electronic Baby Scale
2. Is the item a spare part/attachment or accessory for an existing equipment	NICU & PICU in MCB
3. Name of the manufactures/supplier of the item proposed by the indentor	M/s Rustagi Medisurg
4. Are they sole manufacture/sole distributors of the item	Yes
5. It there any other item with similar/equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, way the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item form these offered by other	NA
6. What were the efforts make to locate alternative source of supply or use other substitutes.	No other make/brand suitable with the existing machine
7. Why open/limited tender can't be resorted to for locating alternative source	Proprietary based item, only manufactured by M/s Rustagi Medisurg. Compatibility of their products.
8. Are the proprietary item certify that the rates are reasonable or not	Yes
9. Any other justification for procuring item from single source.	As above

Signature of Indentor
(Demanding officer)

Rachna Seth

डॉ. रचना सेठ
Dr. RACHNA SETH

I certify that the item at serial no. above is required to be procured on single tender basis as the source of the supply is definitely known, the specified brand proposed was advantage in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Counter signed)
(Head of the Department)

Madhulika Kabra

डॉ. मधुलिका कबरा / Professor & Head

आचार्य एवं विभागाध्यक्ष / Department of Pediatrics

Dr. MADHULIKA KABRA / M.S.

Department of Pediatrics / All India Institute of Medical Sciences

New Delhi-110029

(Strike out whichever is not applicable)

RP*

**DEPARTMENT OF PEDIATRICS
ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

Price Reasonability Certificate

The Department of Pediatrics requires consumable accessory of **ADE Electronic Baby Scale** for the NICU and PICU of Mother & Child Block. This is to certify that M/s Rustagi Medisurg quoted price for the procurement of ADE Electronic Baby Scale for the NICU & PICU of Mother & Child Block amount **Rs. 4,08,870/- (Four Lakhs Eight Thousand Eight Hundred & Seventy Only) including GST** is justified and reasonable.

Rachna Seth

Dr. Rachna Seth
Professor



डॉ. रचना सेठ
Dr. RACHNA SETH
आचार्य/Professor
बालरोग चिकित्सा विभाग/Department of Pediatrics
अ.भा.आ.सं., नई दिल्ली/A.I.I.M.S., New Delhi-29

Madhulika Kabra

Dr. Madhulika Kabra
Professor & Head



डॉ. मधुलिका काबरा/Dr. MADHULIKA KABRA
आचार्य एवं विभागाध्यक्ष/Professor & Head
बाल चिकित्सा विभाग/Deptt. of Pediatrics
अ.भा.आ.सं./A.I.I.M.S.
नई दिल्ली/New Delhi-110029

Infant scales and length measuring devices



Electronic and mechanical baby scales

Electronic baby weighing scale

ADE M118600

The gently contoured baby tray offers generous as well as secure space for small infants and even for seated toddlers.

The fine graduation of 5 g up to 20 kg, together with the auto-hold function and an electronic dampening system guarantees precise and fast weighing results. The clear backlit display ensures easy reading also in dimmed surroundings. Simple to operate by four clearly marked soft touch buttons.

Functions: Unit switch kg/lbs, automatic zero setting, tare, hold, low battery indication, auto switch off, backlight display

CP 20 kg

GR 5 g

UT kg, lbs

TR 600 x 280 x 90 mm

Power supply: Mains and alkaline batteries (4 x AA)



Rustagi Medisurg
87, Darya Ganj,
New Delhi-110002
M-9811028183